भारत सरकार



### GOVERNMENT OF INDIA

ATOMIC ENERGY REGULATORY BOARD

परमाणु ऊर्जा नियामक परिषद्

Chairman

## Case File Number: MH-24145-IR-SU-IRED-E-001

## **Document Number:** 16-COMSUPPTA-112992

**Issuance Date:** 15/06/2016

**Expiry Date:** 15/06/2019

## TYPE APPROVED EQUIPMENT FOR INDUSTRIAL X-RAY EQUIPMENT USED IN INDUSTRIAL RADIOGRAPHY

This has reference to your application no. 16-142027 dated 11/06/2016 . In exercise of the powers conferred under Section-16 of the Atomic Energy Act, 1962 read Rule 3 of the Atomic Energy (Radiation Protection) Rules 2004, the Atomic Energy Regulatory Board (AERB) hereby renews Type Approval from radiological safety view point for the Industrial X-ray Equipment Unit Model Dang Dong XXQ 2505 manufactured by Dan Dong Radiative Instrument Co. Ltd., China with technical specifications given in Annexure-I for sale, installation, operation, servicing and maintenance in India.

## **TERMS AND CONDITIONS**

- 1. This Type Approval becomes invalid if any change is made in Model.
- 2. The Unit shall be supplied only to the user who has obtained requisite permission from AERB to procure the above equipment.
- 3. The supply of the said Model after expiry of this Type Approval is conditional upon its revalidation.
- 4. This Type Approval may be subject to review and modification based on operational experience feedback.
- 5. The additional terms and conditions as specified in Annexure-II.

**Issuing Authority** 

S. A. Bhardwaj Chairman

MR. KULBHUSHAN PREMNATH JETLY EASTWEST ENGINEERING AND ELECTRONICS PVT LTD 44 ACHARYA NAGAR WTMARG, GOVANDI MUMBAI MUMBAI-400088, MAHARASHTRA

Are

परमाणु ऊर्जा नियामक परिषद्,नियामक भवन, अणुशक्तिनगर, मुंबई 400094 (महाराष्ट्र) Atomic Energy Regulatory Board, Niyamak Bhavan, Anushaktinagar, Mumbai 400094 (Maharashtra)



फैक्स/Fax: 91-22-2599 0344

Case File Number: MH-24145-IR-SU-IRED-E-001	<b>Issuance Date:</b> 15/06/2016
Document Number: 16-COMSUPPTA-112992	Expiry Date: 15/06/2019

## ANNEXURE-I

Details	
Type of Equipment	Industrial X-ray Equipment
Model	Dang Dong XXQ 2505
Source	NA
Maximum Activity	NA
Original Equipment Manufacturer	Dan Dong Radiative Instrument Co. Ltd., China
Original Equipment Supplier	Dan Dong Radiative Instrument Co. Ltd., China

# Issuing Authority

S. A. Bhardwaj Chairman

#### ANNEXURE-II

- 1. The supplier shall ensure prior to the supply of the said unit that the user has valid permission from Radiological Safety Division, AERB, Mumbai to procure the unit and the source.
- 2. The unit/model shall be installed only at premises authorised by Radiological Safety Division, AERB, Mumbai.
- 3. The supplier shall be responsible for import, sale, installation, commissioning, servicing and maintenance, supply of spares, source replacement, decommissioning/dismantling of the unit/model and disposal of the radioactive source, as applicable.
- 4. The supplier shall keep AERB informed about the import, sale, installation, commissioning, servicing and maintenance and performance status of the unit/model.
- 5. Point 2 of terms and conditions as mentioned in the type approval certificate is applicable to all equipments except cabinet type x-ray inspection systems.